

Free Water & Particulate Contamination in Distillate Fuels (manual)

ASTM D 4176

Product group(s): Fuel

User group(s): Fuel

Scope: This test method covers two procedures for estimating the presence of suspended free water and solid particulate contamination in distillate fuels having distillation end points below 400 °C and an ASTM color of 5 or less.



Free Water & Particulate Contamination Test Equipment

to be composed of:

- Bar Chart
- Haze-Rating Chart
- Beaker, 1000 ml

Main Unit

13-2350	Bar Chart
13-2351	Haze-Rating Chart
16-0967	Beaker, 1000 ml, high shape without spout

Order Guideline

Minimum equipment: 1x 13-2350, 1x 13-2351, 1x 16-0967
 Spares (approx. 1 year): 16-0967
 Additional requirements: